



OFFICE OF THE CHIEF DISTRICT MEDICAL & PUBLIC HEALTH OFFICER
 CUM DISTRICT MISSION DIRECTOR, ANGUL (NCD CELL)
 (Department of Health & Family Welfare, Govt. of Odisha)
 (npcdcsangul@gmail.com)



Quotation Call Notice No- NCD/ANG/176

Dt-26/02/2026

Quotation call notice for: : Supply of Fluoride Ion Meter.
Date of quotation call notice: : 26/02/2026
Last date and time of submission of quotation: : 10/03/2026 (Till 5 PM)
Date and time of opening of quotation: : 11/03/2026, Time - 11 AM

Sealed quotations are invited from eligible supplier/dealer/manufacture of following items as mentioned below. The supplier / dealer/ manufacture should have valid GST Number and PAN Number. The copy of **GST Number, PAN Number** ^{and catalogue with Specifications of the product} must be attached with the quotation. The quoted amount must **inclusive of all taxes and Other Cost**. The quotation need to be reached in due time and date by Speed post, Regd. Post, or by Courier of the address (O/o the CDM&PHO, Angul, Odisha-759122). The Quotationer should mention clearly in the following format with super-scribing in the envelope as **"Quotation for Supply of Fluoride Ion Meter, CDM & PHO, ANGUL (NCD)"**

Sl.No.	Name of the item	Nos. of item required	Technical Specification for Fluoride Ion meter	Unit Price (quoted price) Inclusive GST												
1	Fluoride Ion Meter	1	<ul style="list-style-type: none"> Large colour graphic LCD display which can show parameter (s) together or each parameter individually. Display should provide additional information of showing reading with connections, time, date, sample ID, User ID. Calibration information, temperature, etc Up to 5 point calibration Meter should have Auto Read continuous display facility ISE measurement should have Auto Blank, Linear Point to Point & Non linear point to point Low Concentration Range Stability. At least 1500 point Data logging with time and date and RS232 for transfer of data to PC Software upgradation facility of equipment through Internet. <p>Technical Specifications</p> <p>ISE Range 0 to 100 ppm Resolution Up to 3 significant digits Units mg/L, ppm, ppb. Power Universal AC Power Adapter, 90-260 VAC, 50-60 Hz</p> <p>ACCESSORIES TO BE SUPPLIED WITH EQUIPMENT</p> <table border="0"> <tr> <td>1. Fluoride Combination, Epoxy Body Electrode</td> <td>-2Nos.</td> </tr> <tr> <td>2 Fluoride Calibration Standard. 1.0 ppm & 100 ppm, 400 ml each</td> <td>-1 No.</td> </tr> <tr> <td>3 TISAB III, at least 400 ml</td> <td>2Nos</td> </tr> <tr> <td>4. TISAB II, at least 3 Liters</td> <td>1 No.</td> </tr> <tr> <td>5 Electrode filling solution, 5 x 60 ml in one packet</td> <td>1No.</td> </tr> <tr> <td>6 Electrode holding Stand</td> <td>1No</td> </tr> </table> <ul style="list-style-type: none"> The Supplier should provide 3 year warranty for the equipment & 2 year AMC Supplier should provide list of users of the equipment in Health Sector during the last five years. 	1. Fluoride Combination, Epoxy Body Electrode	-2Nos.	2 Fluoride Calibration Standard. 1.0 ppm & 100 ppm, 400 ml each	-1 No.	3 TISAB III, at least 400 ml	2Nos	4. TISAB II, at least 3 Liters	1 No.	5 Electrode filling solution, 5 x 60 ml in one packet	1No.	6 Electrode holding Stand	1No	
1. Fluoride Combination, Epoxy Body Electrode	-2Nos.															
2 Fluoride Calibration Standard. 1.0 ppm & 100 ppm, 400 ml each	-1 No.															
3 TISAB III, at least 400 ml	2Nos															
4. TISAB II, at least 3 Liters	1 No.															
5 Electrode filling solution, 5 x 60 ml in one packet	1No.															
6 Electrode holding Stand	1No															

The undersigned has reserved right to reject any or all quotation without assigning any reason thereof. All legal disputes if any relating to the quotation are subject to jurisdiction in the court of law situated at Angul, Odisha only.

Handwritten signature and date: 26.2.26



OFFICE OF THE CHIEF DISTRICT MEDICAL & PUBLIC HEALTH OFFICER
CUM DISTRICT MISSION DIRECTOR, ANGUL (NCD CELL)
(Department of Health & Family Welfare, Govt. of Odisha)
(npcdcsangul@gmail.com)



Memo No: 177 /Date: 26.02.26

Copy to office Notice Board CDM & PHO, Angul and other wings notice board for information and necessary action .


26.2.26
CDM&PHO, Angul

Memo No: 178 /Date: 26.02.26

Copy to all Procurement committee member for information to attend the quotation opening.


26.2.26
CDM&PHO, Angul

Memo No: 179 /Date: 26.02.26

Copy to District E-governance Manager, Collectorate, Angul for information and upload the notice in the District website - www.angul.odisha.gov.in.


26.2.26
CDM&PHO, Angul